


Sehr geehrte Damen und Herren,

Die FFP2 Masken entsprechen der EN149-2001 Norm und haben ein CE-Zeichen.



Packaging: 20PCS/Box

尺寸 : 19.5x9.5x16cm



CE Info:

bsi.

EU Type Examination Certificate

By Royal Charter

This is to certify that:

Jiangyin Changhung Industrial
Manufacturing Factory LLC
Industry Park, No.16, Huannan Road
Huangtang, Xiake Town
Jiangyin
Jiangsu
China

Holds Certificate Number: CE 717693

In respect of:

**Filtering face masks to EN 149:2001+A1:2009,
Various models - see continuation sheets for details.**

on the basis that BSI carried out the relevant Type Examination procedures under the requirements with the Regulation (EU) 2016/425 of the European Parliament and Council relating to Personal Protective Equipment Regulation (PPE) Annex V (Module B) and meets the relevant health and safety requirements specified in Annex II

For and on behalf of BSI, a Notified Body for the above Regulation (Notified Body Number 27977):


Drs. Dave Hagenhaars, Managing Director

Previous Notified Body: BSI 0086

First Issued: 2019-12-06

Latest Issue: 2019-12-06

Effective Date: 2019-12-06

Expiry Date: 2024-12-06

Page: 1 of 7

 **www.grandemask.com** making excellence a habit.

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.
To check its validity telephone +31 (0) 3450790. An electronic certificate can be authenticated [online](#).

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264294, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands
A member of BSI Group of Companies.

EU Type Examination Certificate

No. CE 717693

Product Specification

Flat Fold Vertical Shape

Product Name: Particulate Respirators.
Product Type: Filtering half masks to protect against particles.
Models: CFD3S-P1
CFD3V-P1 (with exhalation valve fitted)

Technical Specification: Harmonized European Standard: EN 149:2001+A1:2009.

Product Description: The particulate respirators are designed to protect against solid and non-volatile liquid particles, are held on the face by a pair of elastic headbands. All models are single shift devices (denoted by the classification symbol NR). These are flat fold vertical shape respirators with valve fitted on one side.

| EN 149 Classification: | Models | Classification |
|------------------------|----------|----------------|
| | CFD3S-P1 | FFP1 NR |
| | CFD3V-P1 | FFP1 NR |

First Issued: 2019-12-06

Latest Issue: 2019-12-06

Effective Date: 2019-12-06

Expiry Date: 2024-12-06

Page: 2 of 7

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.
To check its validity visit www.bsi.nl or +31 20 3460780. An electronic certificate can be authorized [online](http://www.bsi.nl).

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands
A member of BSI Group of Companies.

EU Type Examination Certificate

No. CE 717693

Flat Fold Horizontal Shape

Product Name: Particulate Respirators.
Product Type: Filtering half masks to protect against particles.
Models: CFY3S-P1
CFY3V-P1 (with exhalation valve fitted)
CFY3S-P2
CFY3V-P2 (with exhalation valve fitted)

Technical Specification: Harmonized European Standard: EN 149:2001+A1:2009.

Product Description: The particulate respirators are designed to protect against solid and non-volatile liquid particles, are held on the face by a pair of elastic headbands. All models are single shift devices (denoted by the classification symbol NR). These are flat fold horizontal shape respirators with valve fitted on the top.

| EN 149 Classification: | Models | Classification |
|------------------------|----------|----------------|
| | CFY3S-P1 | FFP1 NR |
| | CFY3V-P1 | FFP1 NR |
| | CFY3S-P2 | FFP2 NR |
| | CFY3V-P2 | FFP2 NR |

First Issued: 2019-12-06

Latest Issue: 2019-12-06

Effective Date: 2019-12-06

Expiry Date: 2024-12-06

Page: 3 of 7

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John H. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.
To check its validity (only for CE 717693), an electronic certificate can be authenticated online.

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John H. Keynesplein 9, 1066 EP Amsterdam, The Netherlands
A member of BSI Group of Companies.

EU Type Examination Certificate

No. CE 717693

Flat Fold Vertical Shape

Product Name: Particulate Respirators.
Product Type: Filtering half masks to protect against particles.
Models: CFB3S-P1
CFB3V-P1 (with exhalation valve fitted)
CFB4S-P1 (with activated carbon)
CFB4V-P1 (with activated carbon and exhalation valve fitted)
CFB3S-P2
CFB3V-P2 (with exhalation valve fitted)
CFB4S-P2 (with activated carbon)
CFB4V-P2 (with activated carbon and exhalation valve fitted)

Technical Specification: Harmonized European Standard: EN 149:2001+A1:2009.

Product Description: The particulate respirators are designed to protect against solid and non-volatile liquid particles, are held on the face by a pair of elastic headbands. All models are single shift devices (denoted by the classification symbol NR). These are flat fold vertical shape respirators with valve fitted on one side.

| EN 149 Classification: | Models | Classification |
|------------------------|----------|----------------|
| | CFB3S-P1 | FFP1 NR |
| | CFB3V-P1 | FFP1 NR |
| | CFB4S-P1 | FFP1 NR |
| | CFB4V-P1 | FFP1 NR |
| | CFB3S-P2 | FFP2 NR |
| | CFB3V-P2 | FFP2 NR |
| | CFB4S-P2 | FFP2 NR |
| | CFB4V-P2 | FFP2 NR |

First Issued: 2019-12-06

Effective Date: 2019-12-06

Latest Issue: 2019-12-06

Expiry Date: 2024-12-06

Page: 4 of 7

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.
To check its validity (telephone: +31 20 5460790), An electronic certificate can be authenticated online.

BSI Group The Netherlands B.V., registered in the Netherlands under number 33254284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands
A member of BSI Group of Companies.

EU Type Examination Certificate

No. CE 717693

Moulded Cup Shape

Product Name: Particulate Respirators.
Product Type: Filtering half masks to protect against particles.
Models: CDW3S-P1
CDW3V-P1 (with exhalation valve fitted)
CDW3S-P2
CDW3V-P2 (with exhalation valve fitted)
CDW-P3 (with exhalation valve fitted)
CDW-P3C (with activated carbon and exhalation valve fitted)

Technical Specification: Harmonized European Standard: EN 149:2001+A1:2009.

Product Description: The particulate respirators are designed to protect against solid and non-volatile liquid particles, are held on the face by a pair of elastic headbands. All models are single shift devices (denoted by the classification symbol NR). These are cup shape moulded style respirators.

| EN 149 Classification: | Models | Classification |
|------------------------|----------|----------------|
| | CDW3S-P1 | FFP1 NR |
| | CDW3V-P1 | FFP1 NR |
| | CDW3S-P2 | FFP2 NR |
| | CDW3V-P2 | FFP2 NR |
| | CDW-P3 | FFP3 NR |
| | CDW-P3C | FFP3 NR |

First Issued: 2019-12-06

Effective Date: 2019-12-06

Latest Issue: 2019-12-06

Expiry Date: 2024-12-06

Page: 5 of 7

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 8, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.
To check its validity contact tel: +31 20 5460780. An electronic certificate can be authenticated online.

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 8, 1066 EP Amsterdam, The Netherlands
A member of BSI Group of Companies.

EU Type Examination Certificate

No. CE 717693

Moulded Cup Shape

Product Name: Particulate Respirators.
Product Type: Filtering half masks to protect against particles.
Models: CDN3S-P1
CDN3V-P1 (with exhalation valve fitted)
CDN4S-P1 (with activated carbon)
CDN4V-P1 (with activated carbon and exhalation valve fitted)
CDN3S-P2
CDN3V-P2 (with exhalation valve fitted)
CDN4S-P2 (with activated carbon)
CDN4V-P2 (with activated carbon and exhalation valve fitted)
CDN-P3 (with exhalation valve fitted)
CDN-P3C (with activated carbon and exhalation valve fitted)

Technical Specification: Harmonized European Standard: EN 149:2001+A1:2009.

Product Description: The particulate respirators are designed to protect against solid and non-volatile liquid particles, are held on the face by a pair of elastic headbands. All models are single shift devices (denoted by the classification symbol NR). These are cup shape moulded style respirators.

| EN 149 Classification: | Models | Classification |
|------------------------|----------|----------------|
| | CDN3S-P1 | FFP1 NR |
| | CDN3V-P1 | FFP1 NR |
| | CDN4S-P1 | FFP1 NR |
| | CDN4V-P1 | FFP1 NR |
| | CDN3S-P2 | FFP2 NR |
| | CDN3V-P2 | FFP2 NR |
| | CDN4S-P2 | FFP2 NR |
| | CDN4V-P2 | FFP2 NR |
| | CDN-P3 | FFP3 NR |
| | CDN-P3C | FFP3 NR |

First Issued: 2019-12-06

Latest Issue: 2019-12-06

Effective Date: 2019-12-06

Expiry Date: 2024-12-06

Page: 6 of 7

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.
To check its validity visit www.bsi.nl or call +31 20 3460780. An electronic certificate can be authenticated [online](http://www.bsi.nl).

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands
A member of BSI Group of Companies.

EU Type Examination Certificate

No. CE 717693

Certificate Administration Details

Technical File Reference:

Certificate Amendment Record

| Issue date | Comments | BSI Review No. |
|---------------|---|-----------------|
| December 2019 | First Issue under PPE Regulation (EU) 425/2016. | 2797:19:3072506 |

Certificate validity

The Certificate holder is responsible for ensuring that the Notified Body is advised of changes to any aspect of the overall process utilised in the manufacture of the product, failure to do so could invalidate the Certificate in respect of product manufactured following the introduction of such changes.

The Conformity to Type Based on Quality Assurance of the Production Process, Annex VIII (Module D) , for the product are referenced in BSI Issued Certificate number CE 717695.

First Issued: 2019-12-06

Latest Issue: 2019-12-06

Effective Date: 2019-12-06

Expiry Date: 2024-12-06

Page: 7 of 7

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.
To check its validity visit www.bsi.nl or call +31 20 360780. An electronic certificate can be authenticated [online](https://www.bsi.nl).

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands
A member of BSI Group of Companies.